

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
 Data File : VX035424.D
 Acq On : 29 Apr 2023 15:20
 Operator : JC/MD
 Sample : 02522-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10D

Quant Time: May 01 04:29:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	248429	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	406010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	145173	48.437	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.880%
35) Dibromofluoromethane	5.385	113	125370	50.310	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.620%
50) Toluene-d8	8.647	98	474730	50.500	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.000%
62) 4-Bromofluorobenzene	11.079	95	178362	50.437	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.880%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3231	2.482	ug/l	98
30) Chloroform	5.092	83	9838	2.059	ug/l	87
64) Tetrachloroethene	9.275	164	9699	3.443	ug/l	97
68) m/p-Xylenes	10.299	106	2288	0.433	ug/l	96
69) o-Xylene	10.640	106	1350	0.261	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
Data File : VX035424.D
Acq On : 29 Apr 2023 15:20
Operator : JC/MD
Sample : 02522-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10D

Quant Time: May 01 04:29:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

