

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
 Data File : VX035425.D
 Acq On : 29 Apr 2023 15:43
 Operator : JC/MD
 Sample : 02522-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: May 01 04:29:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	410474	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	145627	48.947	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	97.900%	
35) Dibromofluoromethane	5.385	113	125261	49.719	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.440%	
50) Toluene-d8	8.647	98	472367	49.702	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	99.400%	
62) 4-Bromofluorobenzene	11.079	95	180711	50.546	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	101.100%	
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	7422	1.565	ug/l	94
64) Tetrachloroethene	9.275	164	954	0.339	ug/l #	85
67) Ethyl Benzene	10.195	91	5670	0.411	ug/l	97
68) m/p-Xylenes	10.305	106	5170	0.979	ug/l	98
69) o-Xylene	10.640	106	1760	0.340	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	3236	0.302	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
Data File : VX035425.D
Acq On : 29 Apr 2023 15:43
Operator : JC/MD
Sample : 02522-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: May 01 04:29:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

