

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
 Data File : VX035415.D
 Acq On : 29 Apr 2023 11:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 01 04:27:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.529	0.492	7.0	88	0.00
3 P	Chloromethane	0.481	0.393	18.3	82	0.00
4 C	Vinyl Chloride	0.483	0.421	12.8#	87	0.00
5 T	Bromomethane	0.321	0.275	14.3	85	0.00
6 T	Chloroethane	0.282	0.247	12.4	88	0.00
7 T	Trichlorofluoromethane	0.755	0.697	7.7	94	0.00
8 T	Diethyl Ether	0.275	0.240	12.7	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.472	6.3	94	0.00
10 T	Methyl Iodide	0.636	0.557	12.4	82	0.00
11 T	Tert butyl alcohol	0.110	0.085	22.7	73	0.00
12 CM	1,1-Dichloroethene	0.486	0.445	8.4#	93	0.00
13 T	Acrolein	0.108	0.105	2.8	94	0.00
14 T	Allyl chloride	0.855	0.751	12.2	87	0.00
15 T	Acrylonitrile	0.275	0.221	19.6	79	0.00
16 T	Acetone	0.262	0.227	13.4	89	0.00
17 T	Carbon Disulfide	1.164	1.041	10.6	87	0.00
18 T	Methyl Acetate	1.013	0.838	17.3	81	0.00
19 T	Methyl tert-butyl Ether	1.666	1.542	7.4	92	0.00
20 T	Methylene Chloride	0.601	0.502	16.5	92	0.00
21 T	trans-1,2-Dichloroethene	0.527	0.473	10.2	91	0.00
22 T	Diisopropyl ether	1.691	1.519	10.2	89	0.00
23 T	Vinyl Acetate	1.162	1.055	9.2	86	0.00
24 P	1,1-Dichloroethane	0.953	0.877	8.0	91	0.00
25 T	2-Butanone	0.382	0.304	20.4	79	0.00
26 T	2,2-Dichloropropane	0.746	0.731	2.0	95	0.00
27 T	cis-1,2-Dichloroethene	0.620	0.572	7.7	94	0.00
28 T	Bromochloromethane	0.406	0.385	5.2	91	0.00
29 T	Tetrahydrofuran	0.245	0.188	23.3	76	0.00
30 C	Chloroform	0.962	0.904	6.0#	93	0.00
31 T	Cyclohexane	0.856	0.764	10.7	88	0.00
32 T	1,1,1-Trichloroethane	0.843	0.810	3.9	94	0.00
33 S	1,2-Dichloroethane-d4	0.603	0.499	17.2	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.307	0.279	9.1	87	0.00
36 T	1,1-Dichloropropene	0.479	0.439	8.4	92	0.00
37 T	Ethyl Acetate	0.438	0.378	13.7	79	0.00
38 T	Carbon Tetrachloride	0.455	0.461	-1.3	94	0.00
39 T	Methylcyclohexane	0.583	0.547	6.2	90	0.00
40 TM	Benzene	1.375	1.268	7.8	90	0.00
41 T	Methacrylonitrile	0.251	0.226	10.0	84	0.00
42 TM	1,2-Dichloroethane	0.490	0.452	7.8	89	0.00
43 T	Isopropyl Acetate	0.735	0.655	10.9	83	0.00
44 TM	Trichloroethene	0.379	0.360	5.0	92	0.00
45 C	1,2-Dichloropropane	0.347	0.330	4.9#	91	0.00
46 T	Dibromomethane	0.238	0.224	5.9	91	0.00
47 T	Bromodichloromethane	0.438	0.439	-0.2	92	0.00
48 T	Methyl methacrylate	0.361	0.331	8.3	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
 Data File : VX035415.D
 Acq On : 29 Apr 2023 11:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 01 04:27:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.007	22.2	76	0.00
50 S	Toluene-d8	1.158	1.006	13.1	82	0.00
51 T	4-Methyl-2-Pentanone	0.452	0.378	16.4	80	0.00
52 CM	Toluene	0.871	0.810	7.0#	89	0.00
53 T	t-1,3-Dichloropropene	0.458	0.482	-5.2	93	0.00
54 T	cis-1,3-Dichloropropene	0.518	0.535	-3.3	92	0.00
55 T	1,1,2-Trichloroethane	0.347	0.323	6.9	90	0.00
56 T	Ethyl methacrylate	0.517	0.489	5.4	88	0.00
57 T	1,3-Dichloropropane	0.594	0.543	8.6	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.249	3.9	87	0.00
59 T	2-Hexanone	0.339	0.285	15.9	78	0.00
60 T	Dibromochloromethane	0.333	0.344	-3.3	93	0.00
61 T	1,2-Dibromoethane	0.358	0.337	5.9	89	0.00
62 S	4-Bromofluorobenzene	0.435	0.378	13.1	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.383	0.364	5.0	93	0.00
65 PM	Chlorobenzene	1.082	1.014	6.3	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.377	-0.3	92	0.00
67 C	Ethyl Benzene	1.873	1.783	4.8#	89	0.00
68 T	m/p-Xylenes	0.718	0.685	4.6	90	0.00
69 T	o-Xylene	0.704	0.675	4.1	90	0.00
70 T	Styrene	1.141	1.115	2.3	90	0.00
71 P	Bromoform	0.262	0.276	-5.3	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.655	3.586	1.9	91	0.00
74 T	N-amyl acetate	1.374	1.298	5.5	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.134	1.024	9.7	85	0.00
76 T	1,2,3-Trichloropropane	0.997	0.867	13.0	87	0.00
77 T	Bromobenzene	0.909	0.878	3.4	89	0.00
78 T	n-propylbenzene	4.198	4.051	3.5	89	0.00
79 T	2-Chlorotoluene	2.623	2.448	6.7	89	0.00
80 T	1,3,5-Trimethylbenzene	3.082	2.991	3.0	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.308	0.298	3.2	85	0.00
82 T	4-Chlorotoluene	2.946	2.829	4.0	91	0.00
83 T	tert-Butylbenzene	3.092	2.981	3.6	91	0.00
84 T	1,2,4-Trimethylbenzene	3.097	3.018	2.6	90	0.00
85 T	sec-Butylbenzene	3.793	3.599	5.1	87	0.00
86 T	p-Isopropyltoluene	3.226	3.135	2.8	89	0.00
87 T	1,3-Dichlorobenzene	1.727	1.660	3.9	91	0.00
88 T	1,4-Dichlorobenzene	1.810	1.646	9.1	90	0.00
89 T	n-Butylbenzene	2.723	2.604	4.4	86	0.00
90 T	Hexachloroethane	0.494	0.510	-3.2	89	0.00
91 T	1,2-Dichlorobenzene	1.727	1.603	7.2	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.216	4.8	83	0.00
93 T	1,2,4-Trichlorobenzene	1.085	1.064	1.9	92	0.00
94 T	Hexachlorobutadiene	0.476	0.439	7.8	89	0.00
95 T	Naphthalene	3.342	3.309	1.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
Data File : VX035415.D
Acq On : 29 Apr 2023 11:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 01 04:27:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.093	1.028	5.9	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6