



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042923\
Data File : VX035414.D
Acq On : 29 Apr 2023 11:13
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Title : SW846 8260
Last Update : Tue Apr 25 16:31:40 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	14713	PASS
75	95	30	60	48.7	43416	PASS
95	95	100	100	100.0	89203	PASS
96	95	5	9	6.8	6064	PASS
173	174	0.00	2	0.4	294	PASS
174	95	50	100	79.2	70645	PASS
175	174	5	9	7.4	5233	PASS
176	174	95	101	97.0	68499	PASS
177	176	5	9	6.2	4223	PASS