

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
 Data File : VX041270.D
 Acq On : 29 Apr 2024 12:16
 Operator : JC/MD
 Sample : VX0429WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBL01

Quant Time: Apr 30 02:06:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	159126	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	229811	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108762	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87582	57.819	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.640%
35) Dibromofluoromethane	5.385	113	72998	52.219	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.440%
50) Toluene-d8	8.647	98	265425	56.277	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	112.560%#
62) 4-Bromofluorobenzene	11.079	95	95081	54.619	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.240%
Target Compounds						
16) Acetone	2.392	43	5779	10.481	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
Data File : VX041270.D
Acq On : 29 Apr 2024 12:16
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Quant Time: Apr 30 02:06:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

