

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
 Data File : VX041276.D
 Acq On : 29 Apr 2024 15:03
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 30 02:10:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147438	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	215304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77625	55.308	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.620%	
35) Dibromofluoromethane	5.385	113	67221	51.327	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.660%	
50) Toluene-d8	8.647	98	245066	55.462	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 110.920%	
62) 4-Bromofluorobenzene	11.079	95	93363	57.245	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.500%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	47617	93.208	ug/l	95
52) Toluene	8.720	92	32313	9.909	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
Data File : VX041276.D
Acq On : 29 Apr 2024 15:03
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 30 02:10:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

