

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
 Data File : VX041278.D
 Acq On : 29 Apr 2024 16:17
 Operator : JC/MD
 Sample : P2261-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RO-RBR200059

Quant Time: Apr 30 02:12:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145581	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	206594	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72892	52.599	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.200%
35) Dibromofluoromethane	5.385	113	62609	49.821	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	8.647	98	231790	54.669	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.340%
62) 4-Bromofluorobenzene	11.079	95	86710	55.407	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.820%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	57385	113.761	ug/l	94
18) Methyl Acetate	2.715	43	1872	1.255	ug/l	93
20) Methylene Chloride	2.788	84	4864	3.948	ug/l #	92
43) Isopropyl Acetate	6.348	43	72465	25.499	ug/l	98
52) Toluene	8.720	92	4795	1.532	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
Data File : VX041278.D
Acq On : 29 Apr 2024 16:17
Operator : JC/MD
Sample : P2261-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RO-RBR200059

Quant Time: Apr 30 02:12:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

