



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042924\
Data File : VX041267.D
Acq On : 29 Apr 2024 08:28
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Title : SW846 8260
Last Update : Tue Apr 23 05:30:28 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	8248	PASS
75	95	30	60	51.5	21995	PASS
95	95	100	100	100.0	42731	PASS
96	95	5	9	6.6	2806	PASS
173	174	0.00	2	0.3	145	PASS
174	95	50	100	98.8	42200	PASS
175	174	5	9	7.6	3215	PASS
176	174	95	101	96.7	40795	PASS
177	176	5	9	6.5	2660	PASS