

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045978.D
 Acq On : 29 Apr 2025 13:35
 Operator : JC/MD
 Sample : Q1901-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB04262025

Quant Time: Apr 30 01:38:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131446	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123412	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66629	54.158	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	5.379	113	48413	51.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	8.647	98	165475	50.834	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.660%	
62) 4-Bromofluorobenzene	11.079	95	63879	53.875	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.740%	
Target Compounds						
16) Acetone	2.386	43	854	1.690	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045978.D
Acq On : 29 Apr 2025 13:35
Operator : JC/MD
Sample : Q1901-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04262025

Quant Time: Apr 30 01:38:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

