

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
 Data File : VX045982.D
 Acq On : 29 Apr 2025 15:08
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 30 01:40:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	62331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126892	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64634	56.703	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.400%	
35) Dibromofluoromethane	5.379	113	46979	52.181	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.360%	
50) Toluene-d8	8.647	98	158707	50.505	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.020%	
62) 4-Bromofluorobenzene	11.079	95	60196	52.590	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042925\
Data File : VX045982.D
Acq On : 29 Apr 2025 15:08
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 30 01:40:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

