

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041295.D
 Acq On : 30 Apr 2024 13:24
 Operator : JC/MD
 Sample : VX0430WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0430WBL01

Quant Time: May 01 02:43:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	199256	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	286230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	267764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	103956	47.924	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.840%
35) Dibromofluoromethane	5.379	113	88541	45.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.300%
50) Toluene-d8	8.647	98	307324	50.058	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.079	95	104097	43.801	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041295.D
Acq On : 30 Apr 2024 13:24
Operator : JC/MD
Sample : VX0430WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0430WBL01

Quant Time: May 01 02:43:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

