

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041299.D
 Acq On : 30 Apr 2024 15:08
 Operator : JC/MD
 Sample : P2265-03DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-02B-14-042424DL

Quant Time: May 01 02:46:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	198623	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	289718	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	107379	49.660	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.320%
35) Dibromofluoromethane	5.385	113	90430	46.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.120%
50) Toluene-d8	8.647	98	329422	53.012	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.020%
62) 4-Bromofluorobenzene	11.079	95	97082	40.358	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.720%#
Target Compounds						
					Qvalue	
21) trans-1,2-Dichloroethene	3.087	96	905	0.526	ug/l	# 71
24) 1,1-Dichloroethane	3.605	63	5108	1.650	ug/l	# 93
27) cis-1,2-Dichloroethene	4.483	96	106080	52.086	ug/l	81
44) Trichloroethene	7.129	130	15783	8.259	ug/l	80
52) Toluene	8.720	92	2015	0.455	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041299.D
Acq On : 30 Apr 2024 15:08
Operator : JC/MD
Sample : P2265-03DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-02B-14-042424DL

Quant Time: May 01 02:46:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

