

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041300.D
 Acq On : 30 Apr 2024 15:31
 Operator : JC/MD
 Sample : P2248-01DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-56-042324DL

Quant Time: May 01 02:46:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	185244	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	269368	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98680	48.933	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.860%
35) Dibromofluoromethane	5.385	113	82713	45.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.620%
50) Toluene-d8	8.647	98	304809	52.757	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.520%
62) 4-Bromofluorobenzene	11.079	95	87059	38.925	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.860%#
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.142	85	4676	3.072	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	244956	168.330	ug/l	# 87
27) cis-1,2-Dichloroethene	4.489	96	10350	5.449	ug/l	80
44) Trichloroethene	7.123	130	657707	370.156	ug/l	87
64) Tetrachloroethene	9.275	164	775	0.484	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041300.D
Acq On : 30 Apr 2024 15:31
Operator : JC/MD
Sample : P2248-01DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-56-042324DL

Quant Time: May 01 02:46:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

