

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041301.D
 Acq On : 30 Apr 2024 15:54
 Operator : JC/MD
 Sample : P2263-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 60156

Quant Time: May 01 02:47:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	186235	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	273813	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98886	48.774	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.540%
35) Dibromofluoromethane	5.385	113	81972	44.177	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.360%
50) Toluene-d8	8.647	98	312001	53.125	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.240%
62) 4-Bromofluorobenzene	11.079	95	113782	50.048	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.100%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	569	0.389	ug/l	Qvalue 92
16) Acetone	2.392	43	45482	47.779	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	2198	0.418	ug/l	97
25) 2-Butanone	4.605	43	3483	2.654	ug/l	# 77
40) Benzene	6.037	78	28448	4.159	ug/l	98
44) Trichloroethene	7.129	130	2885	1.597	ug/l	76
51) 4-Methyl-2-Pentanone	8.580	43	3411	1.366	ug/l	99
52) Toluene	8.720	92	55674	13.306	ug/l	97
67) Ethyl Benzene	10.201	91	4723	0.559	ug/l	98
68) m/p-Xylenes	10.305	106	4607	1.361	ug/l	88
69) o-Xylene	10.640	106	12948	3.996	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	3256	0.523	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041301.D
Acq On : 30 Apr 2024 15:54
Operator : JC/MD
Sample : P2263-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
60156

Quant Time: May 01 02:47:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

