

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041302.D
 Acq On : 30 Apr 2024 16:17
 Operator : JC/MD
 Sample : P2266-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-042324

Quant Time: May 01 02:47:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	178720	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	262175	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96978	49.844	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.680%
35) Dibromofluoromethane	5.385	113	80029	45.044	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.080%
50) Toluene-d8	8.647	98	301161	53.555	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.120%
62) 4-Bromofluorobenzene	11.079	95	87485	40.189	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.380%#
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	4786	2.834	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041302.D
Acq On : 30 Apr 2024 16:17
Operator : JC/MD
Sample : P2266-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-042324

Quant Time: May 01 02:47:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

