

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041305.D
 Acq On : 30 Apr 2024 17:26
 Operator : JC/MD
 Sample : P2248-01DL2 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-56-042324DL2

Quant Time: May 01 02:48:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147262	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	217027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79344	49.492	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.980%
35) Dibromofluoromethane	5.385	113	66760	45.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.780%
50) Toluene-d8	8.647	98	247308	53.128	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.260%
62) 4-Bromofluorobenzene	11.079	95	86971	48.264	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.520%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	37717	32.604	ug/l	87
27) cis-1,2-Dichloroethene	4.495	96	1385	0.917	ug/l	86
44) Trichloroethene	7.123	130	90234	63.031	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041305.D
Acq On : 30 Apr 2024 17:26
Operator : JC/MD
Sample : P2248-01DL2 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-56-042324DL2

Quant Time: May 01 02:48:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

