

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
 Data File : VX041306.D
 Acq On : 30 Apr 2024 17:49
 Operator : JC/MD
 Sample : P2284-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: May 01 02:49:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152613	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	251823	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93472	56.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.520%	
35) Dibromofluoromethane	5.385	113	66247	38.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 77.640%	
50) Toluene-d8	8.647	98	236173	43.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 87.460%#	
62) 4-Bromofluorobenzene	11.079	95	109875	52.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.100%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	6238	4.688	ug/l	98
25) 2-Butanone	4.605	43	3516	3.270	ug/l	98
39) Methylcyclohexane	7.378	83	1937	0.808	ug/l	91
44) Trichloroethene	7.135	130	578	0.348	ug/l	96
68) m/p-Xylenes	10.305	106	1300	0.407	ug/l	100
73) Isopropylbenzene	10.963	105	14425	1.866	ug/l	98
78) n-propylbenzene	11.305	91	25056	3.086	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	20440	3.244	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	42845	6.758	ug/l	94
85) sec-Butylbenzene	11.890	105	28516	3.688	ug/l	87
86) p-Isopropyltoluene	12.006	119	9936	1.431	ug/l	98
89) n-Butylbenzene	12.329	91	29252	5.238	ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043024\
Data File : VX041306.D
Acq On : 30 Apr 2024 17:49
Operator : JC/MD
Sample : P2284-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: May 01 02:49:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

