

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX045991.D
 Acq On : 30 Apr 2025 12:32
 Operator : JC/MD
 Sample : Q1914-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB04282025

Quant Time: May 01 03:36:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	59292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	117758	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60175	55.497	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.000%	
35) Dibromofluoromethane	5.379	113	43965	52.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.240%	
50) Toluene-d8	8.647	98	146264	50.156	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.320%	
62) 4-Bromofluorobenzene	11.079	95	54025	50.860	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX045991.D
Acq On : 30 Apr 2025 12:32
Operator : JC/MD
Sample : Q1914-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB04282025

Quant Time: May 01 03:36:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

