

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX045992.D
 Acq On : 30 Apr 2025 12:56
 Operator : JC/MD
 Sample : Q1904-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-210

Quant Time: May 01 06:50:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	59430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	118869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	114672	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60643	55.799	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.600%
35) Dibromofluoromethane	5.379	113	44462	52.719	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.440%
50) Toluene-d8	8.647	98	152330	51.748	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.500%
62) 4-Bromofluorobenzene	11.079	95	57030	53.187	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.380%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	4135	28.310	ug/l	100
16) Acetone	2.379	43	29047	65.087	ug/l	97
18) Methyl Acetate	2.703	43	1664	1.619	ug/l	99
20) Methylene Chloride	2.788	84	4201	5.028	ug/l	87
25) 2-Butanone	4.562	43	24708	37.823	ug/l	99
43) Isopropyl Acetate	6.348	43	13422	6.171	ug/l	98
59) 2-Hexanone	9.433	43	9886	9.259	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX045992.D
Acq On : 30 Apr 2025 12:56
Operator : JC/MD
Sample : Q1904-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-210

Quant Time: May 01 06:50:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

