

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX045995.D
 Acq On : 30 Apr 2025 14:06
 Operator : JC/MD
 Sample : Q1915-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-04282025

Quant Time: May 01 03:37:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	56553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	111268	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	44562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58476	56.542	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.080%			
35) Dibromofluoromethane	5.379	113	43365	54.931	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.860%			
50) Toluene-d8	8.647	98	146921	53.320	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.640%			
62) 4-Bromofluorobenzene	11.079	95	53005	52.810	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.620%			
Target Compounds						Qvalue
16) Acetone	2.386	43	21932	51.644	ug/l	98
18) Methyl Acetate	2.703	43	1759	1.799	ug/l	96
20) Methylene Chloride	2.788	84	6417	8.071	ug/l	95
43) Isopropyl Acetate	6.348	43	10089	4.956	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	3197	1.079	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX045995.D
Acq On : 30 Apr 2025 14:06
Operator : JC/MD
Sample : Q1915-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-04282025

Quant Time: May 01 03:37:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

