

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX045997.D
 Acq On : 30 Apr 2025 14:53
 Operator : JC/MD
 Sample : Q1905-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-G

Quant Time: May 01 03:37:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122189	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62801	55.632	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.260%
35) Dibromofluoromethane	5.379	113	45023	51.934	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.860%
50) Toluene-d8	8.647	98	156888	51.848	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.700%
62) 4-Bromofluorobenzene	11.079	95	57618	52.276	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.560%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20373	43.951	ug/l	97
18) Methyl Acetate	2.709	43	1834	1.718	ug/l	91
20) Methylene Chloride	2.782	84	3440	3.964	ug/l	91
43) Isopropyl Acetate	6.342	43	11224	5.020	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX045997.D
Acq On : 30 Apr 2025 14:53
Operator : JC/MD
Sample : Q1905-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-G

Quant Time: May 01 03:37:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

