

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX046000.D
 Acq On : 30 Apr 2025 16:03
 Operator : JC/MD
 Sample : Q1906-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-5

Quant Time: May 01 03:38:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	61083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63302	56.669	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.340%	
35) Dibromofluoromethane	5.385	113	44980	51.458	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.920%	
50) Toluene-d8	8.646	98	154369	50.596	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.200%	
62) 4-Bromofluorobenzene	11.079	95	58022	52.210	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.420%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	4955	33.006	ug/l	98
16) Acetone	2.379	43	32947	71.828	ug/l	97
18) Methyl Acetate	2.708	43	2329	2.205	ug/l	93
20) Methylene Chloride	2.788	84	5639	6.567	ug/l	95
25) 2-Butanone	4.568	43	25340	37.740	ug/l	99
43) Isopropyl Acetate	6.348	43	13707	6.081	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX046000.D
Acq On : 30 Apr 2025 16:03
Operator : JC/MD
Sample : Q1906-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-5

Quant Time: May 01 03:38:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

