

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX046001.D
 Acq On : 30 Apr 2025 16:27
 Operator : JC/MD
 Sample : Q1906-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-6

Quant Time: May 01 03:38:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	120942	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62247	55.728	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.460%	
35) Dibromofluoromethane	5.385	113	45253	52.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.480%	
50) Toluene-d8	8.647	98	157197	52.486	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	11.079	95	61140	56.043	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 112.080%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	4961	33.049	ug/l	100
16) Acetone	2.386	43	30185	65.811	ug/l	100
18) Methyl Acetate	2.709	43	2118	2.006	ug/l	96
20) Methylene Chloride	2.782	84	5971	6.954	ug/l	97
25) 2-Butanone	4.568	43	25644	38.196	ug/l	100
43) Isopropyl Acetate	6.342	43	13969	6.312	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX046001.D
Acq On : 30 Apr 2025 16:27
Operator : JC/MD
Sample : Q1906-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-6

Quant Time: May 01 03:38:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

