

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX046002.D
 Acq On : 30 Apr 2025 16:50
 Operator : JC/MD
 Sample : Q1906-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-7

Quant Time: May 01 03:38:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	58345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	114833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	111572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57981	54.341	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.680%			
35) Dibromofluoromethane	5.385	113	42106	51.680	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.360%			
50) Toluene-d8	8.647	98	147959	52.029	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.060%			
62) 4-Bromofluorobenzene	11.079	95	58253	56.237	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 112.480%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	4738	33.042	ug/l	99
16) Acetone	2.386	43	29063	66.334	ug/l	98
18) Methyl Acetate	2.709	43	2568	2.546	ug/l	94
20) Methylene Chloride	2.782	84	5706	6.957	ug/l	93
25) 2-Butanone	4.568	43	25713	40.093	ug/l	96
43) Isopropyl Acetate	6.348	43	13757	6.547	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX046002.D
Acq On : 30 Apr 2025 16:50
Operator : JC/MD
Sample : Q1906-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-7

Quant Time: May 01 03:38:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

