

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX046003.D
 Acq On : 30 Apr 2025 17:13
 Operator : JC/MD
 Sample : Q1907-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CO-8R-WC

Quant Time: May 01 03:38:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/02/2025
 Supervised By :Semsettin Yesilyurt 05/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	120584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61153	54.762	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.520%			
35) Dibromofluoromethane	5.379	113	44293	51.772	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.540%			
50) Toluene-d8	8.647	98	152309	51.005	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.000%			
62) 4-Bromofluorobenzene	11.079	95	56260	51.723	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.440%			
Target Compounds						
16) Acetone	2.380	43	23767	51.831	ug/l	99
18) Methyl Acetate	2.703	43	2312	2.190	ug/l #	90
20) Methylene Chloride	2.782	84	7826	9.116	ug/l	95
25) 2-Butanone	4.587	43	3359m	5.004	ug/l	
43) Isopropyl Acetate	6.361	43	8638	3.915	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX046003.D
Acq On : 30 Apr 2025 17:13
Operator : JC/MD
Sample : Q1907-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CO-8R-WC

Quant Time: May 01 03:38:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/02/2025
Supervised By :Semsettin Yesilyurt 05/05/2025

