

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
 Data File : VX046004.D
 Acq On : 30 Apr 2025 17:37
 Operator : JC/MD
 Sample : Q1912-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-E

Quant Time: May 01 03:38:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123319	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63018	55.712	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.420%
35) Dibromofluoromethane	5.385	113	44278	50.606	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	8.647	98	157527	51.582	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.160%
62) 4-Bromofluorobenzene	11.079	95	58988	53.028	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.060%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	23023	49.568	ug/l	94
18) Methyl Acetate	2.709	43	2163	2.023	ug/l #	86
20) Methylene Chloride	2.782	84	5661	6.510	ug/l	98
43) Isopropyl Acetate	6.348	43	11448	5.074	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX043025\
Data File : VX046004.D
Acq On : 30 Apr 2025 17:37
Operator : JC/MD
Sample : Q1912-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-E

Quant Time: May 01 03:38:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

