


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX050119\
Data File : VX009249.D
Acq On : 01 May 2019 09:17
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X050119W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Thu May 02 06:46:57 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	5885	PASS
75	95	30	60	50.8	14753	PASS
95	95	100	100	100.0	29032	PASS
96	95	5	9	6.6	1914	PASS
173	174	0.00	2	0.3	81	PASS
174	95	50	100	81.8	23757	PASS
175	174	5	9	8.0	1899	PASS
176	174	95	101	96.4	22896	PASS
177	176	5	9	6.8	1547	PASS