

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050120\
 Data File : VX016013.D
 Acq On : 01 May 2020 15:54
 Operator : JC/SP
 Sample : L2470-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 02 07:43:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042820W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 28 17:30:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	288444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	464728	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	434400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206303	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	180776	47.93	ug/l	0.00
Spiked Amount			Recovery	=	95.86%	
35) Dibromofluoromethane	5.47	113	135837	46.86	ug/l	0.00
Spiked Amount			Recovery	=	93.72%	
50) Toluene-d8	8.70	98	562111	51.86	ug/l	0.00
Spiked Amount			Recovery	=	103.72%	
62) 4-Bromofluorobenzene	11.12	95	215582	50.71	ug/l	0.00
Spiked Amount			Recovery	=	101.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050120\
Data File : VX016013.D
Acq On : 01 May 2020 15:54
Operator : JC/SP
Sample : L2470-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 02 07:43:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042820W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 28 17:30:11 2020
Response via : Initial Calibration

