


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX050120\
 Data File : VX016001.D
 Acq On : 01 May 2020 09:25
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X043020W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu Apr 30 12:42:44 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1639

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	11248	PASS
75	95	30	60	51.0	29244	PASS
95	95	100	100	100.0	57389	PASS
96	95	5	9	6.5	3744	PASS
173	174	0.00	2	1.0	476	PASS
174	95	50	100	80.6	46250	PASS
175	174	5	9	7.7	3576	PASS
176	174	95	101	95.6	44205	PASS
177	176	5	9	6.0	2655	PASS