

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
 Data File : VX035438.D
 Acq On : 01 May 2023 14:21
 Operator : JC/MD
 Sample : 02558-08
 Misc : 4.59g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-SS013-1218-01

Quant Time: May 01 23:59:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	242375	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	397711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	357232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163178	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	142754	48.820	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.640%
35) Dibromofluoromethane	5.385	113	110943	45.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.900%
50) Toluene-d8	8.647	98	468854	50.916	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.840%
62) 4-Bromofluorobenzene	11.085	95	169635	48.970	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
Data File : VX035438.D
Acq On : 01 May 2023 14:21
Operator : JC/MD
Sample : 02558-08
Misc : 4.59g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-SS013-1218-01

Quant Time: May 01 23:59:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

