

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
 Data File : VX035442.D
 Acq On : 01 May 2023 15:54
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 02 00:01:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	236049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	384385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	356432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	138109	48.497	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.000%
35) Dibromofluoromethane	5.385	113	113576	48.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	8.647	98	453745	50.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.960%
62) 4-Bromofluorobenzene	11.079	95	178446	53.300	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.600%
Target Compounds						
					Qvalue	
84) 1,2,4-Trimethylbenzene	11.756	105	4052	0.368	ug/l	97
95) Naphthalene	13.780	128	10628	0.895	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
Data File : VX035442.D
Acq On : 01 May 2023 15:54
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 02 00:01:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

