

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
 Data File : VX035444.D
 Acq On : 01 May 2023 16:41
 Operator : JC/MD
 Sample : 02522-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1A

Quant Time: May 02 00:02:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	220767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	365309	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	338323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133183	50.004	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.000%		
35) Dibromofluoromethane	5.385	113	109204	48.705	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.400%		
50) Toluene-d8	8.647	98	430123	50.853	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.700%		
62) 4-Bromofluorobenzene	11.079	95	170553	53.603	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 107.200%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	6806	5.884	ug/l	96
40) Benzene	6.044	78	11279	1.123	ug/l	96
67) Ethyl Benzene	10.195	91	4062	0.320	ug/l	96
68) m/p-Xylenes	10.305	106	2086	0.429	ug/l	89
69) o-Xylene	10.646	106	1187	0.249	ug/l	95
78) n-propylbenzene	11.305	91	2944	0.206	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	13623	1.289	ug/l	97
95) Naphthalene	13.774	128	146573	12.856	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
Data File : VX035444.D
Acq On : 01 May 2023 16:41
Operator : JC/MD
Sample : 02522-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1A

Quant Time: May 02 00:02:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

