

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
 Data File : VX035451.D
 Acq On : 01 May 2023 19:24
 Operator : JC/MD
 Sample : 02515-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-13I

Quant Time: May 02 00:05:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	244135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	392987	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	360168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139348	47.311	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.620%
35) Dibromofluoromethane	5.385	113	120376	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.647	98	457620	50.293	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	175652	51.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.640%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2493	1.949	ug/l	97
95) Naphthalene	13.780	128	13555	1.207	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
Data File : VX035451.D
Acq On : 01 May 2023 19:24
Operator : JC/MD
Sample : 02515-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-13I

Quant Time: May 02 00:05:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

