

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
 Data File : VX035454.D
 Acq On : 01 May 2023 20:34
 Operator : JC/MD
 Sample : 02515-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02

Quant Time: May 02 00:06:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	247759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	396954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	143357	47.960	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.920%
35) Dibromofluoromethane	5.391	113	121473	49.858	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.720%
50) Toluene-d8	8.647	98	467623	50.879	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.760%
62) 4-Bromofluorobenzene	11.079	95	178146	51.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.060%
Target Compounds						
					Qvalue	
84) 1,2,4-Trimethylbenzene	11.750	105	2179	0.204	ug/l	87
95) Naphthalene	13.780	128	12175	1.055	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
Data File : VX035454.D
Acq On : 01 May 2023 20:34
Operator : JC/MD
Sample : 02515-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02

Quant Time: May 02 00:06:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

