

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
 Data File : VX035462.D
 Acq On : 02 May 2023 00:49
 Operator : JC/MD
 Sample : 02521-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-15

Quant Time: May 02 02:44:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	239798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	393259	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	360660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139857	48.343	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.680%
35) Dibromofluoromethane	5.385	113	122266	50.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.300%
50) Toluene-d8	8.647	98	466400	51.223	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.440%
62) 4-Bromofluorobenzene	11.079	95	175927	51.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.720%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12090	9.622	ug/l	97
17) Carbon Disulfide	2.514	76	2333	0.418	ug/l	95
65) Chlorobenzene	10.079	112	90897	11.649	ug/l	100
73) Isopropylbenzene	10.963	105	6799	0.554	ug/l	99
83) tert-Butylbenzene	11.713	119	4450	0.428	ug/l	84
84) 1,2,4-Trimethylbenzene	11.750	105	2374	0.228	ug/l	98
85) sec-Butylbenzene	11.890	105	8681	0.681	ug/l #	67
89) n-Butylbenzene	12.329	91	3467	0.379	ug/l #	83
91) 1,2-Dichlorobenzene	12.335	146	1496	0.258	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	1169	0.321	ug/l #	87
94) Hexachlorobutadiene	13.725	225	722	0.451	ug/l	75
96) 1,2,3-Trichlorobenzene	13.969	180	1030	0.281	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050123\
Data File : VX035462.D
Acq On : 02 May 2023 00:49
Operator : JC/MD
Sample : 02521-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-15

Quant Time: May 02 02:44:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

