

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
 Data File : VX041315.D
 Acq On : 01 May 2024 09:54
 Operator : JC/MD
 Sample : VX0501WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0501WBL01

Quant Time: May 02 03:05:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	179189	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	264066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96332	49.383	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.760%
35) Dibromofluoromethane	5.385	113	77161	43.119	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.240%
50) Toluene-d8	8.647	98	300364	53.031	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.060%
62) 4-Bromofluorobenzene	11.079	95	110596	50.442	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
Data File : VX041315.D
Acq On : 01 May 2024 09:54
Operator : JC/MD
Sample : VX0501WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0501WBL01

Quant Time: May 02 03:05:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

