

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
 Data File : VX041319.D
 Acq On : 01 May 2024 12:25
 Operator : JC/MD
 Sample : P2264-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T11-MW-13-8.0-042324

Quant Time: May 02 03:08:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	188531	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	230473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92145	44.896	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.800%
35) Dibromofluoromethane	5.385	113	82924	53.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.180%
50) Toluene-d8	8.647	98	290099	58.684	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	117.360%#
62) 4-Bromofluorobenzene	11.079	95	122750	64.146	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	128.300%#
Target Compounds						
						Qvalue
16) Acetone	2.392	43	5083	1.740	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	118987	22.342	ug/l	100
20) Methylene Chloride	2.788	84	727	0.408	ug/l	92
31) Cyclohexane	5.464	56	4880	1.970	ug/l	91
39) Methylcyclohexane	7.373	83	6574	2.996	ug/l	95
40) Benzene	6.037	78	4838	0.840	ug/l	# 90
68) m/p-Xylenes	10.311	106	777	0.230	ug/l	77
69) o-Xylene	10.646	106	1308	0.404	ug/l	92
73) Isopropylbenzene	10.963	105	68728	8.334	ug/l	96
78) n-propylbenzene	11.305	91	83956	9.694	ug/l	97
83) tert-Butylbenzene	11.713	119	9738	1.353	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	1967	0.291	ug/l	90
85) sec-Butylbenzene	11.890	105	49132	5.956	ug/l	# 78
89) n-Butylbenzene	12.335	91	13203	2.216	ug/l	# 62
95) Naphthalene	13.774	128	22968	2.481	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
Data File : VX041319.D
Acq On : 01 May 2024 12:25
Operator : JC/MD
Sample : P2264-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-MW-13-8.0-042324

Quant Time: May 02 03:08:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

