

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
 Data File : VX041330.D
 Acq On : 01 May 2024 16:39
 Operator : JC/MD
 Sample : P2264-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-MW-4-9.0-042424

Quant Time: May 02 03:13:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155567	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	229774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91578	54.074	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.140%	
35) Dibromofluoromethane	5.385	113	71418	45.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 91.740%	
50) Toluene-d8	8.647	98	284679	57.763	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 115.520%#	
62) 4-Bromofluorobenzene	11.079	95	109776	57.540	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.080%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	10856	10.814	ug/l	93
78) n-propylbenzene	11.305	91	4712	0.610	ug/l	98
85) sec-Butylbenzene	11.890	105	2269	0.309	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
Data File : VX041330.D
Acq On : 01 May 2024 16:39
Operator : JC/MD
Sample : P2264-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-MW-4-9.0-042424

Quant Time: May 02 03:13:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

