

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
 Data File : VX041332.D
 Acq On : 01 May 2024 17:26
 Operator : JC/MD
 Sample : P2264-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-042324-01

Quant Time: May 02 03:14:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165301	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	259301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97930	54.420	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.840%
35) Dibromofluoromethane	5.379	113	79962	45.505	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
50) Toluene-d8	8.647	98	262744	47.242	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.480%
62) 4-Bromofluorobenzene	11.079	95	95736	44.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.940%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	2148	1.581	ug/l	93
16) Acetone	2.392	43	9152	7.759	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
Data File : VX041332.D
Acq On : 01 May 2024 17:26
Operator : JC/MD
Sample : P2264-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-042324-01

Quant Time: May 02 03:14:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

