


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050124\
Data File : VX041312.D
Acq On : 01 May 2024 08:36
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Title : SW846 8260
Last Update : Wed May 01 02:38:17 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	9304	PASS
75	95	30	60	48.1	25301	PASS
95	95	100	100	100.0	52624	PASS
96	95	5	9	6.7	3532	PASS
173	174	0.00	2	0.8	408	PASS
174	95	50	100	94.5	49707	PASS
175	174	5	9	7.6	3786	PASS
176	174	95	101	98.6	48989	PASS
177	176	5	9	6.7	3292	PASS