

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050219\
 Data File : VX009282.D
 Acq On : 02 May 2019 18:48
 Operator : JC/SP
 Sample : K2659-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW611-190429

Quant Time: May 03 07:52:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299550	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	261884	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112688	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122995	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	5.50	113	91398	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
50) Toluene-d8	8.71	98	374767	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	11.13	95	124407	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	1084	0.477	ug/l	93
16) Acetone	2.45	43	4340	3.195	ug/l	91
21) trans-1,2-Dichloroethene	3.17	96	1079	0.572	ug/l #	81
27) cis-1,2-Dichloroethene	4.60	96	1892	0.841	ug/l	96
52) Toluene	8.79	92	1412	0.270	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
Data File : VX009282.D
Acq On : 02 May 2019 18:48
Operator : JC/SP
Sample : K2659-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW611-190429

Quant Time: May 03 07:52:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

