

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050219\
 Data File : VX009284.D
 Acq On : 02 May 2019 19:35
 Operator : JC/SP
 Sample : K2659-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0516-190429

Quant Time: May 03 03:59:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	186876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	300590	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	263217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125151	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
35) Dibromofluoromethane	5.50	113	92209	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	8.71	98	377731	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	11.13	95	122647	44.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.48%	
Target Compounds						
44) Trichloroethene	7.21	130	3923	1.877	ug/l	97
64) Tetrachloroethene	9.34	164	5336	2.955	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
Data File : VX009284.D
Acq On : 02 May 2019 19:35
Operator : JC/SP
Sample : K2659-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0516-190429

Quant Time: May 03 03:59:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

