

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
 Data File : VX009285.D
 Acq On : 02 May 2019 19:59
 Operator : JC/SP
 Sample : K2659-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW598-190429

Quant Time: May 03 12:15:04 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182175	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	293978	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	253504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104422	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121920	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
35) Dibromofluoromethane	5.50	113	90609	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	8.71	98	368688	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
62) 4-Bromofluorobenzene	11.13	95	118917	43.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.72%	
Target Compounds						
4) Vinyl Chloride	1.40	62	5278	2.362	ug/l #	82
27) cis-1,2-Dichloroethene	4.60	96	2514	1.136	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
Data File : VX009285.D
Acq On : 02 May 2019 19:59
Operator : JC/SP
Sample : K2659-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW598-190429

Quant Time: May 03 12:15:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

