

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050219\
 Data File : VX009287.D
 Acq On : 02 May 2019 20:46
 Operator : JC/SP
 Sample : K2659-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS080MW504-190429

Quant Time: May 03 12:15:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185488	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	302520	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114371	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123902	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
35) Dibromofluoromethane	5.50	113	92169	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	374912	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	11.13	95	125701	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
4) Vinyl Chloride	1.40	62	8993	3.952	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	13357	5.929	ug/l	89
44) Trichloroethene	7.23	130	700	0.333	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
Data File : VX009287.D
Acq On : 02 May 2019 20:46
Operator : JC/SP
Sample : K2659-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS080MW504-190429

Quant Time: May 03 12:15:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

