

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050219\
 Data File : VX009290.D
 Acq On : 02 May 2019 21:56
 Operator : JC/SP
 Sample : K2660-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW052-190429

Quant Time: May 03 07:56:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	183960	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	301057	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	256983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	102594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124576	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
35) Dibromofluoromethane	5.50	113	91460	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	371882	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
62) 4-Bromofluorobenzene	11.13	95	119170	42.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.84%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	703	0.376	ug/l	92
27) cis-1,2-Dichloroethene	4.60	96	6180	2.766	ug/l	92
44) Trichloroethene	7.21	130	6674	3.188	ug/l	97
64) Tetrachloroethene	9.34	164	39416	22.360	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
Data File : VX009290.D
Acq On : 02 May 2019 21:56
Operator : JC/SP
Sample : K2660-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW052-190429

Quant Time: May 03 07:56:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

