

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050219\
 Data File : VX009294.D
 Acq On : 02 May 2019 23:30
 Operator : JC/SP
 Sample : K2660-05DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0512-190429DL

Quant Time: May 03 04:20:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	189300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	306509	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	271001	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	119624	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126683	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
35) Dibromofluoromethane	5.50	113	93385	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	387978	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.13	95	129536	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
16) Acetone	2.44	43	1377	0.992	ug/l	90
44) Trichloroethene	7.22	130	1346	0.632	ug/l	77
64) Tetrachloroethene	9.34	164	32435	17.448	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050219\
Data File : VX009294.D
Acq On : 02 May 2019 23:30
Operator : JC/SP
Sample : K2660-05DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0512-190429DL

Quant Time: May 03 04:20:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

