

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
 Data File : VX028472.D
 Acq On : 02 May 2022 14:53
 Operator : JC/MD
 Sample : N2640-07DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MWBR3-B-117-137-042922DL

Quant Time: May 03 04:50:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	425229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	720824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	704477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	336139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	295582	53.311	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.620%	
35) Dibromofluoromethane	5.391	113	273464	54.097	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.200%	
50) Toluene-d8	8.653	98	1015783	52.318	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.640%	
62) 4-Bromofluorobenzene	11.079	95	369316	49.748	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.500%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	40637	12.435	ug/l	94
24) 1,1-Dichloroethane	3.617	63	7774	1.054	ug/l #	91
27) cis-1,2-Dichloroethene	4.489	96	32529	6.138	ug/l	86
44) Trichloroethene	7.129	130	198944	35.543	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
Data File : VX028472.D
Acq On : 02 May 2022 14:53
Operator : JC/MD
Sample : N2640-07DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MWBR3-B-117-137-042922DL

Quant Time: May 03 04:50:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

