

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
 Data File : VX028478.D
 Acq On : 02 May 2022 17:12
 Operator : JC/MD
 Sample : N2652-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 03 04:52:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	384171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	675896	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	665403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	324396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	286163	57.128	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 114.260%	
35) Dibromofluoromethane	5.391	113	259197	54.683	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.360%	
50) Toluene-d8	8.653	98	970265	53.295	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.600%	
62) 4-Bromofluorobenzene	11.085	95	351942	50.559	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.120%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
Data File : VX028478.D
Acq On : 02 May 2022 17:12
Operator : JC/MD
Sample : N2652-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 03 04:52:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

