

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
 Data File : VX028482.D
 Acq On : 02 May 2022 18:45
 Operator : JC/MD
 Sample : N2658-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40818

Quant Time: May 03 04:53:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	360252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	631812	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	618642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	305347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	261298	55.628	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.260%
35) Dibromofluoromethane	5.391	113	236371	53.347	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.700%
50) Toluene-d8	8.653	98	877504	51.563	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.120%
62) 4-Bromofluorobenzene	11.085	95	324240	49.830	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.660%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	33750	22.252	ug/l	95
20) Methylene Chloride	2.788	84	1854	0.532	ug/l	90
25) 2-Butanone	4.568	43	4386	1.577	ug/l #	85
86) p-Isopropyltoluene	12.012	119	23718	1.253	ug/l #	55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
Data File : VX028482.D
Acq On : 02 May 2022 18:45
Operator : JC/MD
Sample : N2658-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40818

Quant Time: May 03 04:53:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

