

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
 Data File : VX028484.D
 Acq On : 02 May 2022 19:32
 Operator : JC/MD
 Sample : N2606-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3

Quant Time: May 03 04:54:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	360843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	636150	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	662744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	316380	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	263464	55.997	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.000%
35) Dibromofluoromethane	5.391	113	233666	52.377	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.760%
50) Toluene-d8	8.653	98	904412	52.782	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.560%
62) 4-Bromofluorobenzene	11.085	95	336973	51.433	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.860%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12710	7.490	ug/l	97
20) Methylene Chloride	2.794	84	7356	2.107	ug/l #	78
25) 2-Butanone	4.574	43	3931	1.411	ug/l	99
43) Isopropyl Acetate	6.348	43	23792	2.476	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
Data File : VX028484.D
Acq On : 02 May 2022 19:32
Operator : JC/MD
Sample : N2606-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3

Quant Time: May 03 04:54:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

